BD Vacutainer® Safety-Lok™ Blood Collection Set

SAFETY SHIELD ACTIVATION



1 Upon completion of collection, apply light pressure to site using three fingers as shown



3 With the opposite hand, grasp tubing between thumb and index finger



2 Withdraw blood collection set by grasping the translucent yellow safety shield grip area with the thumb and index finger



4 Push the yellow shield forward until the safety shield is locked in place. Discard immediately into an approved sharps container



BD Vacutainer® Safety-Lok™ Blood Collection Set

BD Global Technical Services: 1.800.631.0174

BD Customer Service: 1.888.237.2762

www.bd.com/vacutainer

SAFETY MESSAGE:

Handle all biologic samples and blood collection sharps (lancets, needles, luer adapters, and blood collection sets) in accordance with the policies and procedures of your facility.

Obtain appropriate medical attention in the event of any exposure to biologic samples (for example, through a puncture injury) since they may transmit viral hepatitis, HIV (AIDS), or other infectious diseases.

Utilize a built-in needle protector, if the blood collection device provides one. BD does not recommend reshielding used needles, but the policies and procedures of your facility may differ and should always be followed.

Discard all blood collection sharps in biohazard containers approved for their disposal. Follow your institution's policy for safe disposal of all medical waste.

BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company. ©2004

1 Becton Drive, Franklin Lakes, NJ 07417

Printed in USA 4/04 V55333-3